

typically found, do not pose any

lodine is used as disinfectant in various applications - one of the most common is

the poultry industry waste water treatment. Milwaukee offers MW13 with a range of

0.0 to 12.5 mg/L.

specific health threats to humans. However, excessive contamination of water



Specifications	MW12 Phosphate	MW13 lodine	MW14 Iron
Resolution	0.01 ppm	0.1 ppm	0.01 ppm
Accuracy (@25°C)	±0.04 ppm ±4% of reading	±0.1 ppm ±5% of reading	±0.04 ppm ±2% of reading
Typical EMC Dev.	±0.01 ppm	±0.1 ppm	±0.01 ppm
Light Source	LED @ 525 nm	LED @ 525 nm	LED @ 525 nm
Light Detector	Silicon Photocell	Silicon Photocell	Silicon Photocell
Method	Adaptation of the Standard Methods for the Examination of Water and Wastewater, 20th edition, Ascorbic Acid method. The reaction between phosphate and the reagent causes a blue tint in the sample.	Adaptation of the Standard Methods for the Examination of Water and Wastewater, 18th edition, DPD method. The reaction between iodine and the reagent causes a pink tint in the sample.	Adaptation of the EPA Phenantroline method 315B, for natural and treated waters. The reaction between iron and reagent causes an orange tint in the sample.
Environment	0 to 50°C (32 to 122°F) max 95% RH non-condensing	0 to 50°C (32 to 122°F) max 95% RH non-condensing	0 to 50°C (32 to 122°F) max 95% RH non-condensing
Battery Type	1 x 1.5V AAA	1 x 1.5V AAA	1 x 1.5V AAA
Auto-off	after 2 minutes of non use	after 2 minutes of non use	after 2 minutes of non use
Packaging dimensions	115 x 115 x 84 mm	115 x 115 x 84 mm	115 x 115 x 84 mm
Packaging weight	180 g	180 g	180 g

## **Accessories**

Mi512-25 Phosphate powder reagent (25 pcs) Mi527-25 lodine powder reagent (25 pcs) Mi528-25 Iron powder reagent (25 pcs)

Mi0013

Glass cuvets (2 pcs) Stoppers for cuvets (2 pcs)

## Ordering information:

All handy photometers are supplied in a carton box including 2 cuvets, 6 powder reagents, 1 x 1.5 V AAA battery and instructions

PO<sub>4</sub>/Fe/I

CE